ALL RIGHTS RESERVED. NO PORTION OF THIS PUBLICATION, WHETHER IN WHOLE OR IN PART CAN BE REPRODUCED WITHOUT THE EXPRESS WRITTEN CONSENT OF SPC TECHNOLOGY.

REVISIONS				DOC. NO. SPC-FD04 * Effective: 12/21/98 * DCP No: 680					
DCP #	REV	DESCRIPTION	DRAWN	DATE	CHECKD	DATE	APPRVD	DATE	
		NOT released	MWL	12/30/98					



High frequency replacement probe w/3-postion switchable attenuation, x1, x10 and Ref. 250MHz @ -3dB x 10 position, 10M $\,\Omega$ impedance (1M $\,\Omega$ scope input), 14pF capacitance, <1.4ns risetime. 10MHz @ -3db x 1 position, 1M $\,\Omega$ impedance (scope input), 90pF capacitance, 35ns risetime. Ref position disconnects the probe tip and grounds the scope input. 1.2m cable length with low-noise compensation, 10-35pF, in BNC box. Probe meets EN61010-2-031, 600V (DC+AC peak) CATI P2. Accesssory kit and storage pouch included.

SPC-FOD4.DWG

EXELABILITY ALL STATEMENTS AND TECHNICAL INFORMATION CONTAINED HERDIN ARE BASED UPON INFORMATION AND/OR TESTS WE BELIEVE TO BE ACCURATE AND RELIABLE. SINCE CONDITIONS OF USE ARE BETCHOLO OLAR CONTROL, THE USER SYNAL DESTROAME THE SULTABLITY OF THE PRODUCT FOR THE NIENDED USE AND ASSUME ALL RISK AND LABLITY MHATSOEWOR IN CONNECTION THEREMITH.				TENMA®						
	DRAWN BY:	DATE:	DRAW	DRAWING TITLE:						
	JEFF MCVICKER	12/30/98		250	250 MHz Oscilloscope Probe					
	CHECKED BY:	DATE:	SIZE	DWG. NO. 76-106		ELECTRONIC FILE F				
						84	4N2148.dwg			
	APPROVED BY:	DATE:					T T			
			SCALE	:	U.O.M.:		SHEET: 1 OF	1		